



EC Declaration of Conformity



according to the Medical Devices Directive 93/42/EEC

Class I Medical Device
(non-sterile, non-measuring function)

Manufacturer: Hopefull Medical Equipment Co.,Ltd.

Address: The Yangtze Industrial Park, Economic Development Area, Xian'ning City,
Hubei Province 437100, China

EC Representative: Wellkang Ltd
Suite B, 29 Harley Street, London W1G 9QR, UK

We, the manufacturer, declare under our sole responsibility that

the medical device(s) of class	Product Name	:	Electrical Hospital Bed
	Type/model, identification of product allowing traceability (Where applicable)	:	Please see product list
	according to annex IX of directive 93/42/EEC	:	Class I

is/are in conformity with the relevant provisions and requirements of directive 93/42/EEC, as amended by Directive 2007/47/EC.

Applied harmonised standards, national standards or other normative documents	IEC 60601-2-52:2009	_____
	IEC 60601-1:2005	_____
	EN 60601-1-2:2007	_____

Conformity assessment procedure **Module A (EC Declaration of Conformity (Annex VII) + Technical Files)**

Notified Body (name & number) Certificate & number **NOT applicable**

Signed this Day/ 01 of Month/ 03 of Year/ 2014, Place (Xian'ning City), PR China

Signature (on behalf of the manufacturer) :

Name of authorized signatory: Jianfeng Xie

Position held in the company: General Manager

Official Seal:



Product List

1.Hospital Bed (classification:Class I)

Series Name	Type B series	Type Da series	Type H/Hc series	Type Sa series	Type F series	Children bed	Orthopaedics bed	Transferable bed	Connective bed
Electrical Hospital Bed	B988t-ch B988t B968y Ba868y-ch Ba868y B868y B868y-s B768y B768a B766a B568y B566a	Da858a Da856a Da758a Da756a Da558a Da556a	H838a H538a Hc838L Hc836L Hc738a Hc736a Hc538a Hc536a	Sa738a Sa736a	F868a F538a	Ch738a Ch736a Ch538a Ch536a	Ac768a	NA	NA

Total: 35 Models

HOPEFULL MEDICAL EQUIPMENT CO.,LTD

