



according to the Medical Devices Directive 93/42/EEC Class I Medical Device

(non-sterile, non-measuring function)

Manufacturer:

Hopefull Medical Equipment Co., Ltd.

Address:

The Yangtze Industrial Park, Economic Development Area, Xian'ning City.

Hubei Province 437100. China

EC Representative:

Wellkang Ltd

Suite B. 29 Harley Street, London W1G 9QR, UK

We, the manufacturer, declare under our sole responsibility that

the medical

**Product Name** 

Manual Hospital Bed

**Bedside Accessories** 

device(s)

of class

Type/model, identification of product allowing traceability

Please see product list

(Where applicable)

according to annex IX of directive 93/42/EEC

Class I

is/are in conformity with the relevant provisions and requirements of directive 93/42/EEC, as amended by Directive 2007/47/EC.

Applied harmonised standards, national standards or other normative documents

Comply with the mechanical parts and basic safety performance of the IEC 60601-2-52:2009.

Conformity assessment procedure

Module A (EC Declaration of Conformity (Annex VII) + Technical Files)

Notified Body (name & number) Certificate & number

NOT applicable

Signed this Day/ 01 of Month/ 03 of Year/ 2014 , Place (Xian'ning City ), PR China

Signature (on behalf of the manufacturer):

Name of authorized signatory: Jianfeng Xie EQUIPMENT

Position held in the company: General Manager

Official Seal:

## Product List

1.Hospital Bed (classification:Class I)

Series	Type B series	Type Da series	Type H/Hc series	Type Sa series	Type F series	Children bed	Orthopaedics bed	Transferable bed	Connective bed
Manual Hospital Bed	B668y-m B668y B666a B368y	Da658a Da656a Da358a Da356a Da256a	H338a Hc336a	Sa638a Sa636a Sa338a Sa336a Sa330a Sa236a Sa230a	F338a	Ch638a Ch636a Ch338a Ch336a Ch11 Ch01	Ac638a Ac636a Ac630a	B01	B01-1

2.Air mattress  $\ \ (\mbox{Model:D08},\ \mbox{Classification: Class}\ \ I\ )$ 

3.Movable table (Model:T03, Classification:Class  $\ I\ )$ 

Total: 32 Models

HOPEFULL MEDICAL EQUIPMENT CO.,LTD

March 1st, 2014